

IZJAVA EU O SKLADNOSTI *EU DECLARATION OF CONFORMITY*

(v skladu s prilogo IV Uredbe o medicinskih pripomočkih (EU) 2017/745)
(in accordance with Annex IV of Medical Device Regulation (EU) 2017/745)

Podjetje/Company: **INTERDENT® d.o.o.**
Naslov/Address: **Opekarniška cesta 26, SI - 3000 CELJE**
SRN: **SI-MF-000004584**

Izjavljamo, da smo kot proizvajalec izključno odgovorni za izdajo izjave EU o skladnosti. /
We declare, that as a manufacturer, we are solely responsible for issuing the EU declaration of conformity.

Sledeči proizvodi, razvrščeni v razred I (pravilo 5) po prilogi VIII MDR,
Following Class I Products (rule 5) according to Annex VIII of the MDR,

GENERICNO IME / GENERIC NAME	PRIPOMOČEK ZA DENTALNO FOTOGRAFIJO / ACCESORY FOR DENTAL PHOTOGRAPHY
TRGOVSKO IME / TRADE NAME	PRIPOMOČEK ZA DENTALNO FOTOGRAFIJO / ACCESORY FOR DENTAL PHOTOGRAPHY
GMDN	31776
OSNOVNI UDI-DI / BASIC UDI-DI	++D058DENTALPHOTO14E

ustrezajo splošnim zahtevam glede varnosti in učinkovitosti Uredbe o medicinskih pripomočkih (EU) 2017/745.
comply with general principles and performance requirements of the Medical Devices Regulation (EU) 2017/745.

HARMONIZIRANI IN OSTALI STANDARDI / *HARMONISED AND OTHER STANDARDS:*

EN ISO 13485:2016 Medicinski pripomočki – Sistem vodenja kakovosti – Zahteve za zakonodajne namene /
Medical devices – Quality management systems – Requirements for regulatory purpose

EN ISO 9001:2015- Sistem vodenja kakovosti – zahteve / *Quality management system – Requirements*

EN ISO 14971:2019 Medicinski pripomočki-Uporaba obvladovanja tveganja pri medicinskih pripomočkih /
Medical devices - Application of risk management to medical devices

EN ISO 10993-1:2020 Biološko vrednotenje medicinskih pripomočkov – 1. del: Ocena in preskusi znotraj ocene tveganja / *Biological Evaluation of Medical Devices- Part 1: Evaluation and testing within a risk management process*

EN ISO 15223-1:2021 Medicinski pripomočki – Simboli za označevanje medicinskih pripomočkov - 1. Del: Splošne zahteve / *Medical devices – Symbols to be used with medical device labels, labelling and information to be supplied – Part 1: General requirements*

EN ISO 7405:2018 Zobozdravstvo – Ocena biokompatibilnosti medicinskih pripomočkov v zobozdravstvu /
Dentistry – Evaluation of biocompatibility of medical devices used in dentistry

EN ISO 10993-18:2020 Biološko ovrednotenje medicinskih pripomočkov - 18. del: Kemična opredelitev lastnosti materialov za medicinske pripomočke znotraj procesov obvladovanja tveganja / *Biological evaluation of medical devices – Part 18: Chemical characterization of materials*

ISO/TS 10993-19:2020 Biološko ovrednotenje medicinskih pripomočkov - 19. del: Fizikalno-kemijska, morfološka in topografska karakterizacija materialov / *Biological evaluation of medical devices – Part 19: Physico-chemical, morphological and topographical characterization of materials*

EN 62366-1:2015 Medicinski pripomočki – Del 1: Uporaba inženiringa uporabnosti medicinskih pripomočkov / *Medical devices – Part 1: Application of usability engineering to medical devices*

ISO/TR 20416:2020 Medicinski pripomočki - Nadzor proizvajalcev po dajanju v promet / *Medical devices - Post-market surveillance for manufacturers*

EN ISO 20417:2021 Informacije proizvajalca za medicinske pripomočke / *Information supplied by the manufacturer of medical devices*

EN 1641:2009 Zobozdravstvo - Medicinski pripomočki za zobozdravstvo - Materiali / *Dentistry – Part 1: Medical devices for dentistry - Materials*


Veljavnost izjave o skladnosti je vezana na spremembo medicinskega pripomočka. / *The validity of declaration of conformity is linked to a change in medical device.*

Celje, 16.02.2022

Place, Date

Anja Mavrič, B.Sc.

Responsible person for MD and technical files



Signature:

Verzija / Version: 3